



cicCartuja

centro de
investigaciones científicas
isla de la cartuja

Consejo Superior de Investigaciones Científicas
Junta de Andalucía
Universidad de Sevilla

Specific Use Rules for Calorimetry Service

- ✓ Only authorised users will have access to the equipment. They will be given a user name, a password to access the computer controlling the spectrometer and a secure folder to save their experiment data.
- ✓ To use the equipment, users must reserve on the Biomolecular Interaction Platform web application in advance. Reservations are personal.
- ✓ To guarantee that there are no delays for the next user, make sure everything in the equipment is back in the correct conditions when your shift is finished:
 - Cells are clean and filled with distilled water.
 - Vacuum and cleaning systems are switched off and clean.
 - Syringes are clean and safely stored in their boxes.
 - Plunger head is in its right place.
- ✓ At the end of the shift, users must logout leaving the used material in its right place.
- ✓ Users must register in the log book any incidence or observation that may be relevant to the rest of the users.
- ✓ Authorised users are responsible for their own consumables.
- ✓ Users will be liable for reimbursement if any faults/breakdowns are caused by negligence or misuse of the syringes and equipment or incorrect setup of the experiments.