

Specific Use Rules for DLS Service

- ✓ Only authorised users will have access to the equipment. They Will be given a user name, a password to access the computer controlling the DLS and a secure folder to save their experiment data.
- ✓ To use the equipment, users must reserve on the Biomolecular Interaction Platform web application in advance. Reservations are personal.
- ✓ Turn on the equipment half an hour before starting the analysis.
- ✓ The minimum reserve time is one hour.
- ✓ To ensure that there are no delays for the next user, make sure everything in the equipment is back in the correct conditions (e.g. standard temperature 25 °C) when the shift is finished.
- ✓ At the end of the shift, users must logout leaving and the used material in its right place.
- ✓ Users should communicate any relevant incident or observation directly to the person in charge or by e-mail to bip@ciccartuja.es.
- ✓ Authorised users are responsible for their own consumables.
- ✓ Users will be responsible for refund if any failure/breakdown is caused by negligence or misuse of equipment or incorrect setup of experiments.